

## Technical product specification

<b>Product name</b>	<b>Latex powder free glove</b>	Version / Index no:
<b>Spec code</b>	<b>LCT-059NA-N-3CZ</b>	LCT-059NA-N-3CZ_Version G_January
<b>Date of issue</b>	<b>03.02.2020</b>	2020_EN

### General information

<b>Type</b>	single use examination and disposable protective glove, non sterile
<b>Labelling</b>	information printed on dispenser box
<b>Shape</b>	ambidextrous - straight fingers
<b>Material</b>	Natural Rubber Latex (NRL)
<b>Colour</b>	natural white
<b>Inside</b>	polymer coated / powder free
<b>Outside</b>	no treatment
<b>Cuff / surface</b>	rolled cuff / textured
<b>Shelf life</b>	3 years
<b>Available sizes</b>	XS (5-6) S (6-7) M (7-8) L (8-9) XL (9-10)

### Dimensions, physical properties and biocompatibility

<b>Glove length</b>	median $\geq$ 240 mm (according to EN 455-2)	
<b>Minimum wall thickness</b>	<b>at finger</b>	0.24 mm (double measured) / 0.12 mm (single measured)
	<b>at palm</b>	0.20 mm (double measured) / 0.10 mm (single measured)
	<b>at cuff</b>	0.12 mm (double measured) / 0.06 mm (single measured)
<b>Glove width</b>	according to EN 455-2: median XS $\leq$ 80 mm, S $80 \pm 10$ mm, M $95 \pm 10$ mm, L $110 \pm 10$ mm, XL $\geq 110$ mm	
<b>Force at Break</b>	median $\geq$ 6 N (during shelf life according to EN 455-2)	
<b>Tensile Strength</b>	min. 14 MPa after aging (according to ASTM D3578)	
<b>Elongation at Break</b>	min. 500% after aging (according to ASTM D3578)	
<b>Residual powder / Powder content</b>	$\leq$ 2 mg (according to EN 455-3)	

### Performance requirements and inspection levels

<b>Freedom from holes (Barrier)</b>	AQL $\leq$ 1.5 (as per EN 455-1, sampling in accordance with ISO 2859-1, G-1)
<b>Dimensions and physical properties</b>	AQL 4.0 (as per ASTM D3578, sampling in accordance with ISO 2859-1, S-2)

### Standards, guidelines & quality certificates

<b>Quality certification</b>	ISO 9001, ISO 13485, ISO 14001
<b>Conformity to regulations</b>	Upon request: - Medical Device Regulation (EU) 2017/745: Class I - PPE Regulation (EU) 2016/425: Category I or III - Regulation (EC) 1935/2004 on Food Contact Materials
<b>Conformity to standards</b>	EN 420, EN ISO 374-1, EN 374-2, EN 16523-1, EN 374-4, EN ISO 374-5 (subject to labelling), EN 455-1, EN 455-2, EN 455-3, EN 455-4, ISO 2859-1, ASTM D3578 (except stress at 500% elongation), ASTM F1671

## Technical product specification

<b>Product name</b>	<b>Latex powder free glove</b>	Version / Index no: ·
<b>Spec code</b>	<b>LCT-059NA-N-3CZ</b>	LCT-059NA-N-3CZ_Version G_January
<b>Date of issue</b>	<b>03.02.2020</b>	2020_EN

### Instructions and additional statements

#### Storage instruction

Store in original packaging in a dry and dark place at 10 °C to 30 °C. Refer to guidelines of storage of rubber products as described in ISO 2230:2002. Ensure that storage area is kept cool, dry and dust free, avoid ventilation and storage close to photocopy equipment. Copper ions discolour the glove. Protect gloves against ultraviolet light sources, such as sunlight and oxidizing agents. Storage above 30 °C will lead to accelerated aging and should be avoided.




#### Cautionary statement and ingredient information

This product contains natural rubber latex which may cause allergic reactions, including anaphylactic responses.  
 This product contains accelerators (Dithiocarbamate types, Zinc-mercaptobenzothiazol) not to be used in a hypersensitivity of these substances.  
 For further information, a list of substances contained in the glove is available upon request.

### Reporting system

#### Medical device vigilance and reporting system

According to the official reporting criteria of the Medical Device Regulation, incidents caused by examination gloves must be reported immediately to our Medical Device Reporting team. E-Mail: [sempermed.complaints@semperitgroup.com](mailto:sempermed.complaints@semperitgroup.com) or Tel.: +43 2630 310 0

 A. Wöss Director	 J. Glantschnig   Head of Regulatory Affairs and Contract Management Sempermed	 L. Rieger Head of Product Management
------------------------------------------------------------------------------------------------------------	----------------------------------------------------------------------------------------------------------------------------------------------------------------------------	----------------------------------------------------------------------------------------------------------------------------------

#### Remark

Replaces all previous versions.  
 All standards references refer to the date of document issue.